

Questions & Answers (Set 4)

1. With respect the answer number 2 given in the

"IRIS_IDIQ_RFP_Questions_and_Answers_(Set_3)_Final.pdf it says:

"consistent with the +/-10 degree C operating range without requiring re-centering", however a later requirement "Temperature range over which no mass centering is required" is given as +/- 2 degrees C.

Can you please clarify which is correct?

Answer: The +/-10 degree C operating range without requiring re-centering is correct and applies to Attachment A lines 15, 40, 60 and 85.

2. Moreover, we note that the answer number 10 in the "IRIS_IDIQ_RFP_Pre-Proposal_Conference_Minutes_FINAL-revised.pdf" changes that "+/-2 degree C to +/- 0.05 deg C", which cannot be correct. Please clarify.

Answer: The +/- 0.05 degree C was incorrect. The correct specification should be +/-10 degree C.